


Alpina Plast
Russia, 141603, Moscovskaja Oblast, Klin,
Leningradskoe Shosse, Territoria 88 km,
Korpus 3A, Stroenie 103B.
Tel: +7 495 181 50 42

Place your describtions here for your new test

Statistic for test standard: Default Entry, please modify|500000

| | | | | | | | | |
|----------------|----------|------|--------|--------------------|------------------------|-----|---|--------------------------|
| | Pressure | | Volume | Probe needed Tests | : | 315 | | |
| Average | : | 2.13 | | 6.3 | Valid Tests done | : | 3 | Valid Tests needed : 315 |
| St. Deviation: | : | 0.19 | | 3.3 | Condoms fail Pressure: | : | 0 | Min. for Pressure : 2.0 |
| Minimum | : | 2.00 | | 3.5 | Condoms fail Volume : | : | 3 | Min. for Volume : 18.0 |
| Median | : | 2.35 | | 10.0 | Condoms fail Pin-Hole: | : | 0 | |
| Maximum | : | 2.35 | | 10.0 | Condoms fail Total : | : | 3 | |

DO MORE



| | | | | | | | | | | | | | | |
|-----|------------|----------|----------|-------------|-----|------------|----------|----------|-------------|-----|------------|----------|----------|-------------|
| No | Date | Time | Pressure | Volume S K | No | Date | Time | Pressure | Volume S K | No | Date | Time | Pressure | Volume S K |
| 001 | 2021-07-15 | 17:05:01 | 2.35 | 10.0 - 03 | 002 | 2021-07-15 | 17:05:18 | 2.00 | 5.5 - 03 | 003 | 2021-07-15 | 17:07:00 | 2.05 | 3.5 - 03 |